




POTENTIAL HEALTH EFFECTS OF TOXIC ALUMINUM EXPOSURE¹



Alzheimer's disease and dementia



Congenital heart defects, inflammation and dysfunction of the myocardium, and cardiovascular thrombosis



Pulmonary lesions and pulmonary edema



Crohn's disease and inflammatory bowel diseases



Male/female infertility, pregnancy failures, and poor fetal growth and fetal death



Slow growth and development in children



Osteoporosis, osteomalacia, rickets, exostosis, osteodystrophy, and osteitis fibrosa

IS YOUR BLOOD AND FLUID WARMER UNDER A **CLASS 1 RECALL?**

If you see it listed below, the answer is **YES.**

Smiths Medical
Level 1[®] Fast Flow Warmer

Smisson-Cartledge Biomedical
ThermaCor[®] 1200

Vyaire Medical
enFlow[®] Fluid Warmer

(Recall Terminated on September 22, 2020)

Source:

Potential risk of aluminum leaching with use of certain fluid warmer D. (2021, September 15). U.S. Food and Drug Administration.

<https://www.fda.gov/medical-devices/letters-health-care-providers/potential-risk-aluminum-leaching-use-certain-fluid-warmer-devices-letter-health-care-providers>

701-00480 Rev. A