



## TARGETED TEMPERATURE MANAGEMENT: CritiCool® Infant Step-by-Step

Indication for use: CritiCool is a thermal regulating system, indicated for monitoring and controlling patient temperature

1.	Fill the water tank with sterile water / 0.22 micron filtered water.	
2.	Place the CritiCool unit at the desired location and lock the front wheels.	
3.	Power ON the system. After a short Self-Test, a "Select Mode" screen appears. The water inside CritiCool will begin to cool.	
4.	Select and place the appropriate CureWrap® on the bed or underneath patient (see CureWrap Instructions for Use). <ul style="list-style-type: none"> <li>• <b>Ensure the CureWrap clamps are OPEN.</b></li> <li>• <b>Do not secure the wrap on the patient yet!</b></li> </ul>	
5.	Connect Core cable to the Core socket on the right and Surface cable to the Surface socket on the left of the system, matching colors, gray to gray and green to green, and then plug the temperature probes into the appropriate adapter cables.	
6.	Attach the Core and Surface temperature probes to the patient. <ul style="list-style-type: none"> <li>• <b>Do not place Surface temperature probe underneath the CureWrap.</b></li> <li>• <b>CritiCool cannot thermoregulate without a Core readout.</b></li> </ul>	
7.	Connect water hoses to the wrap and to the CritiCool device. <ul style="list-style-type: none"> <li>• <b>Water automatically starts to flow into the wrap.</b></li> </ul>	
8.	Return to the touch screen. TTM is the default mode and is already highlighted. Touch OK to select TTM.	
9.	The default Set Point (SP) Temperature is 33.5°C. If required, change the Set Point by using the Set Point key: 	
10.	Wait until the CureWrap is filled with water – When filling is complete, secure the CureWrap to the patient (see CureWrap Instructions for Use).	




### Thermoregulation with CritiCool has now begun.

*Note: It is normal to see slight fluctuations in the patient's temperature as a physiological response.*

## CONTROLLED REWARMING

The patient's temperature is gradually increased every hour by a pre-set step.

1.	Choose Select Mode from the drop-down menu on the upper right side of the screen.	
2.	Select <b>Controlled Rewarming</b> mode and touch OK to confirm.	
3.	"Switching to AutoRewarm Mode" message appears. <ul style="list-style-type: none"> <li>• <b>As long as this message appears, the Rewarming has not started yet!</b></li> </ul> Reposition the Core probe if needed, confirm that the Core temperature readout is valid. Wait for the Core reading to stabilize. Touch OK to start the rewarming.	
4.	The patient's temperature will now begin to rise gradually to the required Target Temperature. <ul style="list-style-type: none"> <li>• The CritiCool rewarming Target Temperature default is 36.5°C.</li> <li>• The user can select Target Temperature between 30°C and 40°C using the Target Temperature Key (See user manual instructions).</li> <li>• A Rewarming step (per hour) should be pre-set according to the clinical protocol.</li> <li>• CritiCool uses a default Rewarming step of 0.2°C/hour.</li> </ul>	



## MANUAL REWARMING

The user can manually increase the patient's temperature gradually.

1. The user increases the Set Point temperature manually, as per the rewarming protocol, until the target temperature is achieved.
2. Use Set Point key to select the Set Point.
  - *Note: It is recommended to raise the Set Point temperature in steps of 0.1°C-0.25°C/30 min.*




## MESSAGES

### "Core Readout Too Low"

<b>Trigger:</b>	Core reading is at least $>2.0^{\circ}\text{C}$ below Set Point.
<b>Indicates:</b>	This message may indicate wrong placement of the Core temperature probe.
<b>Impact:</b>	The system will stop the water flow to the CureWrap and halt thermoregulation!

If the message appears:

1. **First** reposition the Core temperature probe if needed.
  2. Check that the Core temperature reading is valid and stable.
  3. **Then** touch OK to confirm Core temperature reading.
- In Controlled Rewarming, it is especially important to ensure the Core temperature reading is correct BEFORE pressing the OK button. If the OK button is pressed before confirming the Core temperature is correct, thermal regulation algorithms may be affected. Thermoregulation will continue and the screen will return to TTM mode. Note: If the system continues to measure low Core temperature, this message will be repeated every 30 minutes.




## MESSAGES

### "Low Core Temperature. Thermoregulation is Continuing..."

<b>Trigger:</b>	If the difference between the patient's Core temperature and the Set Point is $>0.8^{\circ}\text{C}$ but $<2.0^{\circ}\text{C}$ .
<b>Indicates:</b>	This message may indicate wrong placement of the Core temperature probe.
<b>Impact:</b>	Thermoregulation continues.

- If the message appears:
- Check that the Core temperature probe is correctly inserted into the patient. Check that the Core temperature reading is valid and stable.
  - Keep following the patient's temperature.



## AFTER USE CARE & MAINTENANCE

1. Change the system mode to Standby. After waiting a few seconds, close the CureWrap clamps and detach the CureWrap garment and probes from the patient.
2. For cleaning & disinfecting external surfaces, please refer to Chapter 6 of the CritiCool User Manual (DDT136000).
3. For disinfection of the water circuit instructions, please refer to Chapter 6 of the CritiCool User Manual (DDT136000) or the CritiCool water circuit disinfection instructions (701-00653).

\*This document does not replace the requirement to read the entire CritiCool 6.4 user manual.



780 BOSTON ROAD  
BILLERICA, MA 01821, USA  
T: 866.663.0212  
F: 978.663.0214  
WWW.BELMONTMEDTECH.COM

TECHNICAL SUPPORT  
USA: 855.397.4547  
WORLDWIDE: +1.978.663.0212  
TECHSUPPORT@BELMONTMEDTECH.COM

